

# Skeletal Muscle Actin (5C5, F8, C7)

# Mouse anti-human Skeletal Muscle Actin Monoclonal Antibody (Clone5C5, F8, C7)

## **REFERENCES AND PRESENTATIONS<sup>1</sup>**

- ready-to-use (manual or LabVision AutoStainer)
   MAD-005079QD–3 MAD-005079QD–7 MAD-005079QD–12
- Ready-to-use (MD-Stainer)<sup>2</sup> MAD-005079QD-3/V MAD-005079QD/V
- concentrated MAD-005079Q - 1:50 recommended dilution

## COMPOSITION

Anti-human Skeletal Muscle Actin mouse monoclonal antibody purified from serum and prepared in 10mM PBS, pH 7.4, with 0.2% BSA and 0.09% sodium azide **INTENDED USE** To Immunohistochemistry (IHC) on paraffin embedded tissues. Not tested on frozen tissues or Western-Blotting

CLONE: 5C5.F8.C7

Ig ISOTYPE: mouse IgM/k

**IMMUNOGEN:** N-Terminal decapeptide of alpha skeletal muscle isoform of actin; acetylated at the Nterminus.

**SPECIES REACTIVITY:** In vitro diagnostics in humans. Not tested in other species

#### **DESCRIPTION AND APPLICATIONS:**

5C5.F8.C7 MAb is highly specific and shows no crossreaction with smooth muscle actin.

This antibody reacts with sarcomeric actins of normal tissues and neoplasms derived from such tissues (i.e. rhabdomyosarcomas).

IHC POSITIVE CONTROL: Skeletal or Cardiac muscle, or Rhabdomyosarcoma VISUALIZATION: Cytoplasmic

#### **IHC RECOMMENDED PROCEDURE:**

- $4\mu m$  thick section should be taken on charged slides; dry overnight at  $60^{\circ}C$
- Deparaffinise, rehydrate and HIER (heat induced epitope retrieval) – boil tissue in the Pt Module using Vitro S.A EDTA buffer pH8<sup>3</sup> for 20 min at 95°C. Upon completion rinse with 3-5 changes of distilled or deionised water followed by cooling at RT for 20 min
- Endogenous peroxidase block Blocking for 10 minutes at room temperature using peroxidase solution (ref. MAD-021540Q-125)
- Primary antibody: incubate for 20 minutes [The antibody dilution (when concentrated) and protocol may vary depending on the specimen preparation and specific application. Optimal conditions should be determined by the individual laboratory]
- For detection use Master Polymer Plus Detection System (HRP) (DAB included; ref. MAD-000237QK)
- Counterstaining with haematoxylin and final mounting of the slide

**STORAGE AND STABILITY:** Stored at 2-8°C. Do not freeze. Once the packaging has been opened it can be stored until the expiration date of the reagent indicated on the label. If the reagent has been stored under other conditions to those indicated in this document, the user must first check its correct performance taking into account the product warranty is no longer valid.

#### WARNINGS AND PRECAUTIONS:

1. Avoid contact of reagents with eyes and mucous membranes. If reagents come into contact with sensitive areas, wash with copious amounts of water. 2. This product is harmful if swallowed.

3. Consult local or state authorities with regard to recommended method of disposal.

4. Avoid microbial contamination of reagents.

#### SAFETY RECOMMENDATIONS

This product is intended for laboratory professional use only. The product is NOT intended to be used as a drug or for domestic purposes. The current version of the Safety Data Sheet for this product can be downloaded by searching the reference number at <u>www.vitro.bio</u> or can be requested at regulatory@vitro.bio.

For different presentations (references / volumes) please contact the

<sup>1</sup> These references are for presentation in vials of Low Density Polyethylene (LDPE) dropper. In case the products are used in

automated stainers, a special reference is assigned as follows:

/ L: Cylindrical screw-cap vials (QD-3 / L, QD-7 / L, QD-12 / L).
/ N: Polygonal screw-cap vials (QD-3 / N, QD-7 / N, QD-12 / N).



Vitro S.A.

Calle Luís Fuentes Bejarano 60 Ed. Nudo Norte Local 3 41020 Sevilla (Spain) Tel: +34 954 933 200. <u>vitro@vitro.bio</u> ; www.vitro.bio



<sup>3</sup> Ref: MAD-004072R/D

supplier. <sup>2</sup> For Technical specifications for MD-Stainer, please contact your distributor.



#### BIBLIOGRAPHY

1. Schürch W, Skalli O, Seemayer TA, Gabbiani G. Intermediate filament proteins and actin isoforms as markers for soft tissue tumor differentiation and origin. I. Smooth muscle tumors. Am J Path, 128:91-103 (1987).

 Skalli O, Gabbiani G, Babaï F, Seemayer TA, Pizzolato G, Schürch W. Intermediate filament proteins and actin isoforms as markers for soft tissue tumor differentiation and origin. II. Rhabdomyosarcomas. Am J Path, 130:515-531 (1988).
 Cintorino M, Vindigni C, Del Vecchio MT, Tosi P, Frezzotti R, Hadjistilianou T, Leoncini P, Silvestri S, Skalli O, Gabbiani G. Expression of actin isoforms and intermediate filament proteins in childhood orbital rhabdomyosarcomas. J Submicrosc Cytol Pathol, 21:409-419 (1989).

#### LABEL AND BOX SYMBOLS

Explanation of the symbols of the product label and box:

$\square$	Expiration date
Â.	Temperature limit
***	Manufacturer
Σ	Sufficient content for <n> assays</n>
REF	Catalog number
LOT	Lot code
ī	Refer to the instructions of use
IVD	Medical product for <i>in</i> vitro diagnosis.
e-SDS	Material safety data sheet



