

BioCLIA Autoimmune Control Set, PM-Scl

Cat.No.	Kit Size
MY00311	2 X 1 mL
MY00362	4 X 1 mL

INTENDED USE

The BioCLIA Autoimmune Control Set, PM-Scl is intended for the quality control purposes of the BioCLIA PM-Scl performed on the BioCLIA 6500 and BioCLIA 500. For professional *in vitro* diagnostic use only.

SUMMARY AND EXPLANATION

PM/Scl overlap syndrome or scleromyositis is a complex autoimmune disease (a disease in which the immune system attacks the body). Patients with scleromyositis have symptoms of both systemic scleroderma and either polymyositis or dermatomyositis, and is therefore considered an overlap syndrome. Although it is a rare disease, it is one of the most common overlap syndromes seen in scleroderma patients, together with MCTD and Antisynthetase syndrome. Autoantibodies often found in these patients are the anti-PM/Scl (anti-exosome) antibodies.¹

The symptoms that are seen most often are typical symptoms of the individual autoimmune diseases and include Raynaud's phenomenon, arthritis, myositis and scleroderma.² Treatment of these patients is therefore strongly dependent on the exact symptoms with which a patient reports to a physician and is similar to treatment for the individual autoimmune disease, often involving either immunosuppressive or immunomodulating drugs.^{3,4}

MATERIALS SUPPLIED

- PM-Scl Control N** Barcode labeled tubes with buffer containing human antibodies to PM-Scl in stabilizers and preservatives. Ready to use, 1 mL.

Control	N
---------	---

Preservatives: 0.0015% < Proclin 300 < 0.6%.

- PM-Scl Control P** Barcode labeled tubes with buffer containing human antibodies to PM-Scl in stabilizers and preservatives. Ready to use, 1 mL.

Control	P
---------	---

Preservatives: 0.0015% < Proclin 300 < 0.6%.

The Control Code contains controls' information is provided in each kit.

Target value and acceptance range for the controls are indicated on the card provided in each kit.

WARNINGS AND PRECAUTIONS

- For professional *in vitro* diagnostic use only.
- Do not use any controls beyond their expiration dates.
- Do not mix controls from different lots unless specified.
- Instructions must be carefully followed for using and storing of

controls. Any modification in procedure may interfere with the results.

- Controls and contaminated vials must be handled strictly following safety guidelines or rules of biological hazards to ensure the users' and environmental safety.
- Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established!

Precautions:

Human serum is added in the controls.

- The human derived material in this product was tested by FDA approved methods and found nonreactive for Hepatitis B Surface Antigen (HBsAg), Anti-HCV and HIV 1/2 antibodies. Handle as if potentially infectious. 7 Avoid contacting with skin and eyes. Do not empty into drains. Wear suitable protective clothing.



- Proclin 300 is added in the controls at concentration between 0.0015% - 0.6%.
- Controls contain chemical and biological components. Avoid ingesting or splashing onto skin and mucous membrane. If direct contact with controls happens, rinse the contact surface with plenty of water immediately and see a doctor if necessary.

STORAGE CONDITIONS

- Store the kit at 2-8 °C.
- The shelf life of the unopened kit is 12 months from day of production.
- Vial opened controls could be used for 28 successive days, exposure no more than 2 hours each time when kept uncapped and is good for up to 35 controls, after which the reagent must be discarded.
- Avoid repeated freezing and thawing.

ASSAY PROCEDURE

Detailed information about operating the BioCLIA instruments can be taken from the Instrument User's Manual.

Note that, it is important to perform all routine maintenance procedures for optimal performance.

Control

The control procedure should be done before running the specimens each day. Users also can adjust the control procedure period according to their own lab frequency.

Each Laboratory should establish its own reference ranges.

Programming and Running samples

1. Put the kit into the corresponding position of the reagent chamber of the fully automatic chemiluminescence analyzer. The information of the kit can be uploaded into the instrument system through the scanning of reagent barcode, and can also be set through the supporting software of the instrument.
2. The information of calibrator / quality control is identified by scanning the calibrator / control barcodes, and the position of calibrator / quality control is assigned in the instrument system.
3. The sample to be tested is placed on the instrument sample rack chamber, and the corresponding test information is edited through the instrument supporting software.
4. Start the operation procedure, and all calibrator / quality control / sample processing steps will be automatically executed.

TRACEABILITY

The reported values were determined over multiple runs on the BioCLIA 6500 and BioCLIA 500 using specific lots of reagents against an in-house standard. DGP IgA results are reported in RU/mL which is interpreted from relative light unit (RLU). Method comparison test showed good sensitivity and specificity of tested assay.

LIMITATIONS

- The controls are designed for control of the same lot of BioCLIA Autoimmune Reagent Kit.
- The controls can be kept uncapped onboard the instrument up to 2 hours for each time of usage. And a total up to 35 controls are suggested, for any longer period of time, the reagent should be discarded, otherwise may result in improper results.

SYMBOLS

	Catalog Number		Use-by date
	In Vitro diagnostic medical device		Lot Number
	Store between +2°C and +8°C		Consult Instruction for Use
	Manufacturer		Authorized Representative in the European Community
	CE Marking		Contains Sufficient for <n> Tests
	Biological Risk		GHS07 Warning

	Negative Control
	Positive Control

REFERENCE

1. Jablonska S, Blaszczyk M. Scleroderma Overlap Syndromes. Advances in Experimental Medicine & Biology 1999;455:85-92.
2. Mahler M, Raijmakers R. Novel aspects of autoantibodies to the PM/ScI complex:

clinical, genetic and diagnostic insights. Autoimmunity Reviews 2007;6:432-7.

3. Jablonska S, Blaszczyk M. Scleromyositis: a scleroderma/polymyositis overlap syndrome. Clinical Rheumatology 1998;17:90-91.

4. Vanderghenst F, Ocmant A, Sordet C, Humbel RL, Goetz J, Roufosse F, et al. Anti-pm/scl antibodies in connective tissue disease: Clinical and biological assessment of 14 patients. Clinical & Experimental Rheumatology 2006;24:129-33.

5. Richmond JY, Mckinney RW. Biosafety in microbiological and biomedical laboratories: U.S.GPO.1999.



HOB Biotech Group Corp., Ltd.
C6 Building, No. 218 Xinghu Road, Suzhou Industrial Park,
Suzhou, Jiangsu, 215123, China

CONTACT INFORMATION:


TEL (+86)512-69561996

Fax (+86)512-62956652

WEBSITE: www.hob-biotech.com

CUSTOMER SERVICE: HOB Biotech Group Co., Ltd

CUSTOMER SERVICE: TEL (+86)4008601202



EUROPE REPRESENTATIVE: Emergo Europe

ADDRESS/LOCATION:

Prinsessegracht 20, 2514 AP The Hague, The Netherlands



The eIFU is available on Website:

<http://en.hob-biotech.com/usercenter/login.aspx>

TECHNICAL ASSISTANCE

For technical assistance, contact your National Distributor.

Date of issue: 17th March 2019

Date of revision: 20th December 2021

Change Control Number: CN21129E

Version: A/1 (EN)