
Doc #: PI9284 Version #: 2



# MSH<sub>2</sub> **Control Slides**







# **Intended Use**

For In Vitro Diagnostic Use.

## **Summary and Explanation**

MSH2 is a mismatch repair gene commonly associated with Hereditary Non-Polyposis Colorectal Cancer. This gene was identified as a locus frequently mutated in Hereditary Non-Polyposis Colorectal Cancer. When cloned, it is a human homolog of the E. coli DNA mismatch repair gene mutS, consistent with the characteristic alterations in microsatellite sequences (RER+ phenotype) found in Hereditary Non-Polyposis Colorectal Cancer.

MSH2 is abnormally deficient in a high proportion of patients with microsatellite instability (MSI-H). This finding is associated with the autosomal dominant condition found in Hereditary Non-Polyposis Colon Cancer. This anti-MSH2 antibody (along with MLH1 antibody) is useful in screening patients and families for this rare condition. Colon cancers that are microsatellite unstable have a better prognosis than their microsatellite stable counterparts.

# Presentation

Five slides of MSH2 positive tissues, each mounted on Hydrophilic Plus Slides, provided in a plastic mailer.

Catalog No.	Quantity
BSB-9284-CS	5 slides
BSB 5777	5 slides
BSB-3722-CS	5 slides

# Storage Store at 20-25°C

#### **Precautions**

- 1. For professional users only. Results should be interpreted by a qualified medical professional.
- 2. Ensure proper handling procedures are used with this reagent.
- 3. Always wear personal protective equipment such as a laboratory coat, goggles, and gloves when handling reagents.
- 4. Dispose of unused solution with copious amounts of water.
- 5. Follow safety precautions of the heating device used for epitope retrieval (TintoRetriever Pressure Cooker or similar).
- 8. For additional safety information, refer to Safety Data Sheet for this product.
- 9. For complete recommendations for handling biological specimens, please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (see References in this document).

## Stability

This product is stable up to the expiration date on the product label. Do not use after expiration date listed on package label.

## **IHC Protocol**

- 1. Subject tissues to heat induced epitope retrieval (HIER) using a suitable retrieval solution such as ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023) or EDTA (BSB 0030-BSB 0033).
- 2. Any of three heating methods may be used:

# a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin iar containing the ImmunoDNA Retriever with Citrate or EDTA and place on trivet in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high. Incubate for 15 minutes. Open and immediately transfer slides to room temperature.

# b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 30-60 minutes.

# c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a steamer. cover and steam for 30-60 minutes.

- 3. After heat treatment, transfer slides in ImmunoDNA Retriever with Citrate or EDTA to room temperature and let stand for 15-20 minutes.
- 4. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 5. Wash slides with ImmunoDNA washer or DI water.
- 6. Continue IHC staining protocol. Wash slides between each step with ImmunoDNA washer solution.

### Abbreviated Immunohistochemical Protocol

Abbreviated illimationistochemical i rotocot			
Step	ImmunoDetector AP/HRP	PolyDetector AP/HRP	PolyDetector Plus HRP
Peroxidase/AP Blocker	5 min.	5 min.	5 min
Primary Antibody	30-60 min.	30-60 min.	30-60 min.
1st Step Detection	10 min.	30-45 min.	15 min.
2nd Step Detection	10 min.	Not Applicable	15 min.
Substrate- Chromogen	5-10 min.	5-10 min.	5-10 min.
Counterstain / Coverslip	Varies	Varies	Varies

#### Abbreviated IF Protocol

7.55141444 11 11040404		
Step	Incubation Time	
Rinse slides in IF wash buffer	5 minutes	
Drain and wipe excess IF wash buffer off slide		
Conduct remaining steps in the dark		
Apply Antibody	30-60 minutes	
Rinse with 3 changes of IF wash buffer	3x15 minutes each	
Coverslip with IF mounting medium		

# **Mounting Protocols**

For detailed instructions using biodegradable permanent mounting media such as XyGreen PermaMounter (BSB 0169-0174) or organic solvent based resin such as PermaMounter (BSB 0094-0097), refer to PI0174 or PI0097.

## **Product Limitations**

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized, and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a qualified medical professional.

#### References

- 1. Wright CL, et al. Am J Surg Pathol. 2003;27:1393-1406
- 2. Brueckl WM, et al. Anticancer Research. 2003:23:1773-1778
- 3. Rigau V, et al. Arch Pathol Lab Med. 2003;127:June:694-700
- 4. Renkonen E, et al. J Clin Oncol. 2003;21:3629-3637
- 5. Hoedema R, et al. The American Surgeon. 2003; May 69(5): 387-92
- 6. U.S. Department of Health and Human Services: Centers for Disease Control and Prevention. Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories. Supplement / Vol. 61, January 6, 2012

https://www.cdc.gov/mmwr/pdf/other/su6101.pdf

Symbol Key / Légende des symboles/Erläuterung der Symbole





