


Trop-2/EGP-1 Control Slides





Intended Use

For In Vitro Diagnostic Use.

Summary and Explanation

Trop-2 or Tumor-associated calcium signal transducer 2, also known as epithelial glycoprotein-1 antigen (EGP-1), is a protein that in humans is encoded by the TACSTD2 gene. This transmembrane glycoprotein functions in a variety of cell signaling pathways and was first elucidated as a transducer of an intracellular calcium signal. Trop-2 expression has been demonstrated to depend on a large variety of transcription factors. Trop-2 is involved in several cell signaling pathways, of which many are associated with tumorigenesis. In Thyroid Cancer cell invasion, Trop-2 signal transduction has been seen as a downstream effect of the ERK and JNK pathways, where its signaling enhances stem cell-like properties of cancer cells, as Trop-2 regulates proliferation and self-renewal through ■-catenin signaling. In a study, the majority of Papillary Thyroid Carcinoma specimens were positive for Trop-2; however, the pattern of staining differed significantly between the histopathological variants. All Papillary microcarcinomas, PTC classic variants, and tall cell variants were Trop-2 positive, with mainly diffuse staining. In contrast, less than half of the PTC follicular variant specimens were positive for Trop-2, with only focal immunoreactivity. Trop-2 may play a role in tumor progression given the involvement in several molecular pathways traditionally associated with cancer development. High Trop-2 expression correlates with poor prognosis in Pancreatic Carcinoma, Hilar Cholangiocarcinoma, Cervical Cancer, Gastric Cancer, and others. Increased Trop-2 expression has been associated with poor overall and disease-free survival outcomes across several solid tumors. Given Trop-2's expression pattern and associated poor prognostic outcomes, Trop-2 is a rational prognostic marker and a possible therapeutic target.

Presentation

Five slides of Trop-2/EGP-1 positive tissues, each mounted on Hydrophilic Plus Slides, provided in a plastic mailer.

Catalog No.	Quantity		
BSB-9419-CS	5 slides		
BSB-3754-CS	5 slides		

Storage Store at 20-25°C

Precautions

- 1. For professional users only. Results should be interpreted by a qualified medical professional.
- 2. Ensure proper handling procedures are used with this reagent.
- 3. Always wear personal protective equipment such as a laboratory coat, goggles, and gloves when handling reagents.
- 4. Dispose of unused solution with copious amounts of water.
- 5. Follow safety precautions of the heating device used for epitope retrieval (TintoRetriever Pressure Cooker or similar).
- 8. For additional safety information, refer to Safety Data Sheet for this product.
- 9. For complete recommendations for handling biological specimens, please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (see References in this document).

Stability

This product is stable up to the expiration date on the product label. Do not use after expiration date listed on package label.

IHC Protocol

- 1. Subject tissues to heat induced epitope retrieval (HIER) using a suitable retrieval solution such as ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023) or EDTA (BSB 0030-BSB 0033).
- 2. Any of three heating methods may be used:

a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA and place on trivet in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high. Incubate for 15 minutes. Open and immediately transfer slides to room temperature.

b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 30-60 minutes.

c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a steamer, cover and steam for 30-60 minutes.

- 3. After heat treatment, transfer slides in ImmunoDNA Retriever with Citrate or EDTA to room temperature and let stand for 15-20 minutes.
- 4. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 5. Wash slides with ImmunoDNA washer or DI water.
- 6. Continue IHC staining protocol. Wash slides between each step with ImmunoDNA washer solution.

Abbreviated Immunohistochemical Protocol

Step	ImmunoDetector AP/HRP	PolyDetector AP/HRP	PolyDetector Plus HRP					
Peroxidase/AP Blocker	5 min.	5 min.	5 min					
Primary Antibody	30-60 min.	30-60 min.	30-60 min.					
1st Step Detection	10 min.	30-45 min.	15 min.					
2nd Step Detection	10 min.	Not Applicable	15 min.					
Substrate- Chromogen	5-10 min.	5-10 min.	5-10 min.					
Counterstain / Coverslip	Varies	Varies	Varies					

Abbreviated IF Protocol

71551 CVINICUIT 1 1 1 0 1 0 1 0 1					
Step	Incubation Time				
Rinse slides in IF wash buffer	5 minutes				
Drain and wipe excess IF wash buffer off slide					
Conduct remaining steps in the dark					
Apply Antibody	30-60 minutes				
Rinse with 3 changes of IF wash buffer	3x15 minutes each				
Coverslip with IF mounting medium					

Mounting Protocols

For detailed instructions using biodegradable permanent mounting media such as XyGreen PermaMounter (BSB 0169-0174) or organic solvent based resin such as PermaMounter (BSB 0094-0097), refer to PI0174 or PI0097.

Product Limitations

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized, and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a qualified medical professional.

References

1. Ripani E, Sacchetti A, Corda D, Alberti S. Human Trop-2 is a tumor-associated calcium signal transducer. Int J Cancer. 1998;76(5):671-676.

doi:10.1002/(sici)1097-0215(19980529)76:5<671::aid-ijc10>3.0.co;2-7 2. Guan H, Guo Z, Liang W, et al. Trop2 enhances invasion of thyroid cancer by inducing MMP2 through ERK and JNK pathways. BMC Cancer. 2017;17(1):486. Published 2017 Jul 14. doi:10.1186/s12885-017-3475-2

3. Stoyanova T, Goldstein AS, Cai H, Drake JM, Huang J, Witte ON. Regulated proteolysis of Trop2 drives epithelial hyperplasia and stem cell self-renewal via β -catenin signaling. Genes Dev. 2012;26(20):2271-2285. doi:10.1101/gad.196451.112

4. Fong D, Moser P, Krammel C, et al. High expression of TROP2 correlates with poor prognosis in pancreatic cancer. Br J Cancer. 2008;99(8):1290-1295. doi:10.1038/sj.bjc.6604677

5. Liu T, Liu Y, Bao X, Tian J, Liu Y, Yang X. Overexpression of TROP2 predicts poor prognosis of patients with cervical cancer and promotes the proliferation and invasion of cervical cancer cells by regulating ERK signaling pathway. PLoS One. 2013;8(9):e75864. Published 2013 Sep 27. doi:10.1371/journal.pone.0075864

6. Zeng P, Chen MB, Zhou LN, Tang M, Liu CY, Lu PH. Impact of TROP2 expression on prognosis in solid tumors: A Systematic Review and Meta-analysis. Sci Rep. 2016;6:33658. Published 2016 Sep 20. doi:10.1038/srep33658

7. Bychkov A, Sampatanukul P, Shuangshoti S, Keelawat S. TROP-2 immunohistochemistry: a highly accurate method in the differential diagnosis of papillary thyroid carcinoma. Pathology. 2016;48(5):425-433. doi:10.1016/j.pathol.2016.04.002

8. U.S. Department of Health and Human Services: Centers for Disease Control and Prevention. Guidelines for Safe WorkPractices in Human and Animal Medical Diagnostic Laboratories. Supplement / Vol. 61, January 6, 2012. https://www.cdc.gov/mmwr/pdf/other/su6101.pdf

Symbol Key / Légende des symboles/Erläuterung der Symbole

EC REI	QAdvis EAR AB Ideon Science Park Scheelevägen 17 SE-223 70 Lund, Sweden	1	Storage Temperature Limites de température Zulässiger Temperaturbereich	***	Manufacturer Fabricant Hersteller	REF	Catalog Number Référence du catalogue Bestellnummer
IVD	In Vitro Diagnostic Medical Device Dispositif médical de diagnostic in vitro In-Vitro-Diagnostikum	[] i	Read Instructions for Use Consulter les instructions d'utilisation Gebrauchsanweisung beachten	\subseteq	Expiration Date Utiliser jusque Verwendbar bis	LOT	Lot Number Code du lot Chargenbezeichnung

