


# Cyclin D3 Control Slides





## **Intended Use**

For In Vitro Diagnostic Use.

## **Summary and Explanation**

Cyclin D3 is a protein that in humans is encoded by the CCND3 gene. Cyclin D3 forms a complex with and functions as a regulatory subunit of CDK4 or CDK6, required for the G1/S cell cycle phase transition. Abnormal expression of Cyclin D3 was found in many different cancers and may be an unfavorable prognostic biomarker for various malignancies.

Expression of Cyclin D3 was significantly associated with and shown to be an independent prognostic marker in Breast Cancer; high Cyclin D3 expression indicates poor disease free survival rate in Breast Cancer. Increased p27 expression in Hurthle Cell variant of Follicular Thyroid Carcinoma was shown to be caused by the overexpression of Cyclin D3. Cyclin D3 overexpression also plays a role in the progression of Pancreatic Adenocarcinoma, as well as androgen receptor-dependent proliferation in Prostatic Adenocarcinoma.

## Presentation

Five slides of Cyclin D3 positive tissues, each mounted on Hydrophilic Plus Slides, provided in a plastic mailer.

Catalog No.	Quantity
BSB-9437-CS	5 slides

Storage Store at 20-25°C

#### **Precautions**

- 1. For professional users only. Results should be interpreted by a qualified medical professional.
- 2. Ensure proper handling procedures are used with this reagent.
- 3. Always wear personal protective equipment such as a laboratory coat, goggles, and gloves when handling reagents.
- 4. Dispose of unused solution with copious amounts of water.
- 5. Follow safety precautions of the heating device used for epitope retrieval (TintoRetriever Pressure Cooker or similar).
- 8. For additional safety information, refer to the Safety Data Sheet for this product.
- 9. For complete recommendations for handling biological specimens, please refer to the CDC document, "Guidelines for Safe Work Practices in Human and Animal Medical Diagnostic Laboratories" (see References in this document).

## **Stability**

This product is stable up to the expiration date on the product label. Do not use after expiration date listed on package label.

#### **IHC Protocol**

- 1. Subject tissues to heat induced epitope retrieval (HIER) using a suitable retrieval solution such as ImmunoDNA Retriever with Citrate (BSB 0020-BSB 0023) or EDTA (BSB 0030-BSB 0033).
- 2. Any of three heating methods may be used:

## a. TintoRetriever Pressure Cooker or Equivalent

Place tissues/slides in a staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA and place on trivet in the pressure cooker. Add 1-2 inches of distilled water to the pressure cooker and turn heat to high. Incubate for 15 minutes. Open and immediately transfer slides to room temperature.

#### b. TintoRetriever PT Module or Water Bath Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA at 95°-99° C. Incubate for 30-60 minutes.

## c. Conventional Steamer Method

Place tissues/slides in a pre-warmed staining dish or coplin jar containing the ImmunoDNA Retriever with Citrate or EDTA in a steamer, cover and steam for 30-60 minutes.

- 3. After heat treatment, transfer slides in ImmunoDNA Retriever with Citrate or EDTA to room temperature and let stand for 15-20 minutes.
- 4. For manual staining, perform antibody incubation at ambient temperature. For automated staining methods, perform antibody incubation according to instrument manufacturer's instructions.
- 5. Wash slides with ImmunoDNA washer or DI water.
- 6. Continue IHC staining protocol. Wash slides between each step with ImmunoDNA washer solution.

#### Abbreviated Immunohistochemical Protocol

Step	ImmunoDetector AP/HRP	PolyDetector AP/HRP	PolyDetector Plus HRP
Peroxidase/AP Blocker	5 min.	5 min.	5 min
Primary Antibody	30-60 min.	30-60 min.	30-60 min.
1st Step Detection	10 min.	30-45 min.	15 min.
2nd Step Detection	10 min.	Not Applicable	15 min.
Substrate- Chromogen	5-10 min.	5-10 min.	5-10 min.
Counterstain/Coverslip	Varies	Varies	Varies

### Abbreviated IF Protocol

Step	Incubation Time	
Rinse slides in IF wash buffer	5 minutes	
Drain and wipe excess IF wash buffer off slide		
Conduct remaining steps in the dark		
Apply Antibody	30-60 minutes	
Rinse with 3 changes of IF wash buffer	3x15 minutes each	
Coverslip with IF mounting medium		

## **Mounting Protocols**

For detailed instructions using biodegradable permanent mounting media such as XyGreen PermaMounter (BSB 0169-0174) or organic solvent based resin such as PermaMounter (BSB 0094-0097), refer to PI0174 or PI0097.

## **Product Limitations**

Due to inherent variability present in immunohistochemical procedures (including fixation time of tissues, dilution factor of antibody, retrieval method utilized, and incubation time), optimal performance should be established through the use of positive and negative controls. Results should be interpreted by a qualified medical professional.

#### References

- 1. Chi Y, et al. Cyclin D3 predicts disease-free survival in breast cancer. Cancer Cell Int. 2015;15:89.
- 2. Troncone G, et al. Accumulation of p27(kip1) is associated with cyclin D3 overexpression in the oxyphilic (Hurthle cell) variant of follicular thyroid carcinoma. J Clin Pathol. 2007;60(4):377-381.
- 3. Ito Y, et al. Expression and possible role of cyclin D3 in human pancreatic adenocarcinoma. Anticancer Res. 2001;21(2A):1043-1048.
- 4. Olshavsky NA, et al. Cyclin D3 action in androgen receptor regulation and prostate cancer. Oncogene. 2008;27(22):3111-3121.

Symbol Key / Légende des symboles/Erläuterung der Symbole

OAdvis EAR AB Storage Temperature Manufacturer Catalog Number Ideon Science Park EC **REP** REF Limites de température Fabricant Référence du catalogue Scheelevägen 17 Zulässiger Temperaturbereich Hersteller Bestellnummer SE-223 70 Lund, Sweden Read Instructions for Use In Vitro Diagnostic Medical Device **Expiration Date** Lot Number **IVD** Consulter les instructions  $\begin{bmatrix} \mathbf{i} \end{bmatrix}$ LOT Dispositif médical de diagnostic in vitro Utiliser jusque Code du lot d'utilisation In-Vitro-Diagnostikum Verwendbar bis Chargenbezeichnung Gebrauchsanweisung beachten

